

Addendum

The competent authority has approved following changes in the technical specifications:

Tender Clause Ref:	Existing	May be read as
Section VII	Integrated ELSD and PDA variable wavelength UV detector with automatic collection of fractions at all available wavelengths from 200 to 400 nm or more.	Integrated ELSD and PDA variable wavelength UV detector with automatic collection of fractions at all available wavelengths from 200 to 800 nm or more.
	Fraction collector with different size racks.	Fraction collector with different size racks with 18 ml & 25 ml tubes.
	Safety features like over pressure sensor, lighted rack area, audible alarm & optional feature of fraction collector rack enclosure.	Safety features like over pressure sensor, lighted rack area, audible alarm and with feature of fraction collector rack enclosure.
	Must quote the consumable for at least for one year to run the system with full capacity. (Cost to be included in the main Quotation only, the price of units should be mentioned separately).	120 number 4gm normal phase column, 40 number 12gm normal phase column and 10 number 40gm normal phase column, 20 number empty loading cartridges, 10 number reusable empty columns and 3 stainless steel reusable column. All these columns should be provided along with the machine.
	Quoted model must have at least 12 installations in Delhi NCR, and the bidders should submit a installations report of the quoted model. Bidder should also provide a user list and at least 5 user satisfactory certificates from recognized research Institute/University in India where instrument quoted model has been extensively used for similar applications.	Quoted model must have at least 25 installations in India, and the bidders should submit installations report of the quoted model. Bidder should also provide a user list and at least 5 user satisfactory certificates from recognized Govt. research Institute/University in India where instrument quoted model has been extensively used for similar applications
	Additional Specification	Any breakdown call must be attended within 24 hours.

-SD XXX-

Tender Inviting Authority