

# Corrigendum

**Tender No. RCB/ATN/13/15-16/DAILAB (Fully Motorized Upright Epifluorescence Microscope For Automated Widefield Stitching, With Scientific CMOS Camera And High End Image Analysis Work Station).**

The following specifications have been modified:

1. Specification no. 5: Widefield ergonomic trinocular phototube with 3-way light path distribution for simultaneous observation and imaging, and 2 camera ports.
2. Specification no. 6: Motorized Z-drive with Z-step size of at least 25 nm or better, having coarse and fine drive knob preferably with detachable/remote control option to reduce vibrations. A nose-piece motorized Z focus mechanism with fixed stage is preferred.
3. Specification no. 13: Objectives of the following specifications should be quoted: 2X/2.5X (NA 0.08 or better), Plan Achromat 20X (NA 0.75 or better) with DIC, Plan Achromat 40X (NA 0.80 or better) with DIC, and Plan Achromat 60X/63X oil (NA 0.135 or better)".

All other specifications and the terms and conditions of the tender remain unchanged.