

CORRIGENDUM-2

Corrigendum-2 for Tender File No. RCB/ATN/30/17-18. (Automated Stereozoom Microscope with Fluorescence Imaging)

Subsequent to the uploading of corrigendum-1 based on the pre-bid meeting discussions, the specifications of the aforementioned tender for the stereomicroscope with fluorescence, were further reviewed and the following additional changes are made in this tender.:

- Added requirement for 4 or more position fluorescent turret along with band pass fluorescence filters for GFP and RFP (Specification SI.No.16).
- Microscope should be able to achieve high depth and high resolution simultaneously (Specification SI.No.18).
- System should be upgradable in future for maximum possible optical attachments (Specification SI.No.19).