

### Corrigendum Pre Bid Clarifications

**Tender No:** RCB/NIT/10/20-21/CIS

**Tender Name:** Supply installation and commissioning of Chemidoc Imaging System along with Image analysis software

S. No.	Name of the Agency	Bidding Document Reference(s) (Section number/page number)	Content of RFP requiring clarification	Clarification raised by the bidder	RCB Remarks
1	Internal Scientific suggestion/discussion	Technical Specification Point No.20 General Specifications	Additional point General Specifications	One 4-gel vertical SDS gel running apparatus should be provided.	Point accepted
2	Internal Scientific suggestion/discussion	Technical Specification Point No..13 system software	Additional Point System software	For chemiluminescence application, software should be able to capture protein ladder (molecular weight marker control) along with the protein of interest at the same time.	Point accepted


Further, bidding deadline may be please extended by one week.



(Dr. Rajender Mehta)



(Dr Prasad Ahare)



(Dr. Ramu Subram)