

Corrigendum Pre Bid Clarifications

Tender No: RCB/NIT/13/20-21/HSCIS

Tender Name: Supply, installation and commissioning of High Speed Live Cell Radiometric Calcium Imaging Station.

S. No.	Name of the Agency	Bidding Document Reference(s) (Section number/page number)	Content of RFP requiring clarification	Clarification raised by the bidder	RCB Remarks
1	Nikon (Towa) and Olympus (DSS)	Technical Specification Point No.13	High sensitive sCMOS camera.	Addition of Dynamic Range of Camera (1:36000 or better). Other specs of Camera remain same,	Point accepted
2	Nikon (Towa) and Olympus (DSS)	Technical Specification Point No.14	Monitor size of the computer.	Increase size of monitor to 27". Other specs of Computer and related accessories remain same.	Point accepted
3	Nikon (Towa), Olympus (DSS) and Leica (DHR)	Technical Specification Point No. 19. a.	Optional: Suitable Anti-vibration table	Optional: Suitable Anti-vibration table with active air dumping/air-compression technology.	Point accepted

Nikhil
13/20-21/HSCIS

[Signature]

[Signature]
SAM MATHEW

Dr. Praveen Anave

(Dr. Tejendra Motiani)