

**Post Pre-bid clarification / Corrigendum**

The points raised by prospective bidders during the pre-bid meeting were discussed in detail and following decisions were taken: -

Tender Ref.	Item / Equipment	Original point in tender	Amendment / New addition
RCB/NIT/15/19-20/BBB/CC	Cell Counter	System should have the facility to use one-time usable slides for performing cell counting, imaging, and other analysis.	System must have the facility to use both re-usable slides and one-time disposable slides for performing cell counting, imaging and other analysis.
		At least 100 one-time usable slides (along with cover slips, if required) must be included within the quoted price of the instrument as standard. If a secondary feature of re-usable slide/cytometer is available with the system, then it is mandatory that such featured item be included within the quoted price for the instrument as standard – in such case, quoting 100 one-time usable slides can be ignored.	System must come with 2 units of compatible re-usable slide (along with min. 20 coverslips, if required) as well as 1 box of one-time disposable slides (min. 50 nos.).
		Price of 500 one-time usable cell counting slides must be quoted as an accessory option to procure <b><u>and the same will be compared during financial bid evaluation as well</u></b> – this compliance must also be followed by system with secondary feature of re-usable slide/cytometer.	<b>*Removed</b>
		NOTE: The aforementioned asked price of 500 one-time usable slides will be fixed for the next 5 years but will be considered for procurement <b><u>only if required</u></b> . And if required, the bidder as well as the OEM must accept such	<b>*Removed</b>

	order of consumable slides (500 nos.) for 'n' number of times in the next five years.	
	Credibility: The quoted model must be from a proven international manufacturers of automated cell counters with installations worldwide. They must show a comprehensive list of at least 10 installations of the quoted model or a similar model in reputed Indian research laboratories in the last 5 years. (Installation reports must be included with dates signified).	Credibility: The quoted model must be from a proven international manufacturers of automated cell counters with installations worldwide. They must show a comprehensive list of at least 10 installations of the quoted model or a similar model, preferably in reputed Indian research laboratories in the last 5 years. (Installation reports must be included with dates signified).
	<p>GUARANTEE/WARRANTEE PERIOD: THE TENDERERS MUST PROVIDE FULL SYSTEM WARRANTY FOR 5 YEARS FROM THE DATE OF COMPLETION OF THE SATISFACTORY INSTALLATION FOLLOWED BY A COMPREHENSIVE AMC WARRANTY QUOTE (RATES) <b><u>DIRECTLY FROM THE OEM</u></b> FOR THE SUBSEQUENT 5 YEARS WHICH MUST INCLUDE ALL SPARES, ACCESSORIES AND LABOR. FAILURE TO COMPLY WITH THIS CONDITION WILL ENTAIL THE REJECTION OF THE BIDS. THE PRICE COMPARISION SHALL BE MADE TAKING INTO ACCOUNT THE BASIC PRICE AND POST WARRANTY CMC.</p>	<p>The bidders must quote full system warranty for 5 years from the OEM. Post-warranty comprehensive maintenance contract for another 5 years must be separately quoted which must include all spares, accessories and labor. THE PRICE COMPARISON SHALL BE MADE TAKING INTO ACCOUNT THE BASIC PRICE AND POST WARRANTY CMC. FAILURE TO COMPLY WITH THIS CONDITION WILL ENTAIL THE REJECTION OF THE BIDS.</p>

**\*Each and every technical specification required must be supported by original company catalogue with respectively highlighted point. If not highlighted in technical brochure/catalogue, the bid will not be allowed.**