

Corrigendum Pre Bid Clarifications

Tender No: RCB/NIT/15/20-21/CO2O2

Tender Name: Supply, installation and commissioning of CO2+O2 Temperature Analyzer.

S. No.	Name of the Agency	Bidding Document Reference(s) (Section number/page number)	Content of RFP requiring clarification	Clarification raised by the bidder	RCB Remarks
1	Eppendorf India	Technical Specification Point No.4	System should have CO2 Measuring accuracy of $\pm 0.2\%$	Measuring accuracy of $\pm 1\%$ instead of $\pm 0.2\%$	Point accepted
2	Eppendorf India	Technical Specification Point No.6	System should have O2 Measuring accuracy of $\pm 0.2\%$	Measuring accuracy of $\pm 1\%$ instead of $\pm 0.2\%$	Point accepted

[Signature]
19/11/2020

SAM MATHEW

[Signature]

Dr Prasad Abnave

Dr. Rajendar Motani