

Corrigendum

Supply, installation and commissioning of Probe Sonicator for bacterial cell disruption with sound proof enclosure

NIT No: RCB/NIT/25/19-20/PS

The probes should be made from lead free material to avoid leaching/mixing of lead with the cell-lysates/expressed proteins which can obstruct the experiments. The claim should be substantiated with documentation such as company brochure/certification.

