

# Regional Centre for Biotechnology

an institution of education, training & research

(Established by the Dept of Biotechnology, Govt of India under the auspices of UNESCO)

180, Udyog Vihar Phase I, Gurgaon (Haryana) – 122016

Phone: 0124-4264103; email: rcb@nii.ac.in

No. RCB/NIQ/Freezer/02/14-15

Date : 11 Sep 2014

As per attached list in Annexure-III

## SUBJECT: NOTICE INVITING QUOTATIONS

We are interested in the purchase of the articles mentioned below/as per list enclosed. Please send your quotation in a sealed cover and indicate (i) Name of the materials (ii) Reference of this letter and (iii) date of opening of Quotation. The quotations must accompany a Demand Draft amounting to Rs. 5,000/- being the EMD in the name of Executive Director RCB, Gurgaon. Your **SEALED** quotation should reach this office **on or before 10.10.2014 upto 15:00 hrs. The same will be opened on 10.10.2014 at 15:30 hrs.**

Sl. No	Description	Quantity
1	4° Celsius Upright Refrigerators for Laboratory use ( Specifications as per Annexure-I )	01
2	-20°C Upright Deep Freezer for Laboratory use ( Specifications as per Annexure-II )	01

N.B.:

1. While submitting quotations please take note of the instructions overleaf. Quotations are liable to be rejected if any of these conditions are not complied with. Please note that vendor may quote for single item also. RCB reserves the right to place order of all the items collectively or separate order for each item.
2. The quantity shown above are just indicative. The same will be increased/decreased to any extent depending upon actual requirement.
3. Orders will be placed with successful bidder(s) for each item as and when required basis. The price should be kept valid for one year from the date of bid opening.

Prices quoted should be for:-

- ( a ) Free delivery at RCB Gurgaon.
- ( b ) F.O.R. Gurgaon by Passenger train / Road Transport / Air India / Indian Airlines.
- ( c ) Insurance ware house to ware house.
- ( d ) Freight.

**Adm. Offr.**

## **TERMS AND CONDITIONS**

- (1) We are interested in the material either of indigenous manufacturer or of foreign make available from ready stock. Any offer to supply on Forward Delivery Basis under suppliers own quota licence will also be considered.
- (2) Your rates should include packing and forwarding charges. The goods should be insured against theft loss or breakage during transit and insurance charges.
- (3) The rates of sales tax should be clearly indicated wherever chargeable.
- (4) Specific mention should be made of whether the delivery will be ex-stock or will have to be imported and how much time will be required for delivery after placing of the order.
- (5) In case of supply order for the SCIENTIFIC EQUIPMENT / APPARATUS the date. of delivery should be strictly adhered to otherwise the supply order is liable to be cancelled.
- (6) In the case of supply order for stores other than Scientific Equipment/Apparatus: - As time is essence of this order, the date of delivery should be strictly adhered to otherwise the Executive Director, RCB reserves the right not to accept delivery in the part or full and claim Liquidated damages 1 % per week subject to maximum of 10% of the total value of supply order.
- (7) In case any of the items mentioned overleaf is on the current rate contract list, Please quote D.G.S. & D. rate and rate contract number.
- (8) The quotation will be opened in the presence of the representatives/in presence of officers of RCB, on the date mentioned in main document..
- (9) The payments will be made in 30 days by Account Payee Cheque after receipt of the materials in good condition.
- (10) The acceptance of the quotation will rest with the Executive Director, RCB who does not bind himself to accept to lowest quotation and reserves the right to himself to reject or partially accept any or all the quotations received without assigning any reason.

### **IMPORTANT NOTES:-**

- (1) If you are on D.G.S. & D. rate contract, please quote D.G.S. & D. rate & rate contract no., enclosing a copy thereof.
- (2) For an offer of imported material, please give full break up of your rates FOB/Insurances/Pkgs & Forwarding/ Freight of your Principal's invoice/Quotation, as, the case may be.
- (3) Please indicate your Permanent Income Tax Number (PAN Number) on your Performa Invoice/Bill.
- (4) Printed conditions of the quotation shall not be binding on us.
- (5) Quotations must be clearly written or typed without any cutting or over writing.
- (6) Dispute, if any, will be Subject to Jurisdiction of Delhi/NCR.
- (7) Please quote your standard Warranty.

**SPECIFICATIONS FOR 4° CELSIUS UPRIGHT REFRIGERATORS FOR LABORATORY USE**

1. Unit should be specifically designed/meant for Laboratory/Research centers/Hospitals/Clinics and for scientist use.
2. Gross /Net volume of 370 liters, with 1850x595x634 mm (HxWxD) approximate dimensions and not more than 90 kg in weight.
3. 4\*star or greater rating and T climate class with A/A+ energy class.
4. Temperature range from +1° C to +8° C with auto defrost system.
5. Energy consumption of approximately 0.39/0.48 kWh/24hr.
6. Should use R600A as refrigerant and C-Pentane as blowing agent for insulation.
7. Refrigerator should be compatible for data logger for recording temperature parameters.
8. Refrigerator should have static refrigeration system and condenser type, with natural/forced type of airflow.
9. Refrigerator should have a single door, with option for reversibility with lock and key.
10. Refrigerator should have rear rollers and leveling feet.
11. Should have adjustable shelves, specifically meant for laboratory use.
12. Should maintain uniform temperature in the cabinet.
13. Should be endorsed/certified with highest quality standards.
14. Should be quoted with suitable stabilizer.
15. Compatible microprocessor controller /audio visual alarm system should be quoted as optional items.
16. Credibility: Manufacturers should demonstrate a record of successful installations in similar reputed institutions/ research centers across India. A comprehensive list of prior installations should be provided.
17. Two years comprehensive warranty for entire system.
18. The post warranty CMC for another three years should be quoted separately. The price comparison shall be taken into account with the basic cost of the equipment with 2 years warranty and post warranty CMC for three years. Failure to comply this condition entails rejection of bids.
19. Satisfactory performance certificate at least from three users where the same model is installed should be submitted.
20. The selected vendor should also agree to relocate and re-install the equipment at Faridabad Biotech Cluster without any additional financial implication.

**SPECIFICATIONS FOR -20 °C UPRIGHT DEEP FREEZER FOR  
LABORATORY USE**

1. Unit should specially designed/meant for laboratory / Research centres /Hospital /Clinical and for the scientist use.
2. Gross /Net volume of 340-350 litres, with 24” x 25” x 73” (WxDxH) approximate dimensions and not more than 90 kg in weight.
3. 4\* star or greater rating and T climate class with A+ / A energy class.
4. Temperature range of – 16 °C to – 25 °C temperature.
5. Energy consumption of approximately 1.77 units / 0.94 kWh / 24hr.
6. Should use R 600 A as refrigerant and C - Pentane as blowing agent for insulation.
7. Freezer should be compatible for data logger for recording temperature parameters.
8. Freezer should have static refrigeration system and condenser type, with natural / forced type of airflow.
9. Freezer should have single door with option for reversibility and lock and key.
10. Freezer should have rear rollers and levelling feet.
11. Should have adjustable shelves.
12. Should maintain uniform temperature in the cabinet.
13. Should be endorsed/certified with highest quality standards.
14. Compatible microprocessor controller /audio visual alarm system should be quoted as optional items for freezer.
15. Should have quick freezing function.
16. Should be suitable for fixing optional temperature alarms.
17. Should be frost free.
18. Optional facility for fixing the alarms like door opening/ acoustic.
19. Should be quoted with suitable stabilizer.
20. Credibility: Manufacturers must demonstrate a record of successful installations in similar reputed institutions/research centers across India. A comprehensive list of prior installations should be provided.
21. Two years comprehensive warranty for entire system.
22. The post warranty CMC for another three years should be quoted separately. The price comparison shall be taken into account with the basic cost of the equipment with 2 years warranty and post warranty CMC for three years. Failure to comply this condition entails rejection of bids.
23. Satisfactory performance certificate at least from three users where the same model is installed should be submitted.
24. The selected vendor should also agree to relocate and re-install the equipment at Faridabad Biotech Cluster without any additional financial implication.

**List of Firms to whom Enquiry Sent**

M/s.DOLCHE INDIA,  
C-32, DLF, Sector-14,  
Gurgaon – 122 001.

M/S Maruti Electronics  
A-50 Devli Extn  
Tigri Road  
New Delhi-110062

M/S Shubh Electronics  
C-207, Sangam Vihar  
New Delhi-110062

M/s Scientific Systems,  
B-5, Ashoka Chambers,  
Flat No. 3A, Rajindra Park,  
Pusa Road,  
New Delhi - 60

M/s Hitech Associates Pvt. Ltd  
Shop No. 32 Sector 15-G,  
DDA Market, Rohni  
Delhi-110043

M/s Scientific Systems,  
B-5, Ashoka Chambers,  
Flat No. 3A, Rajindra Park,  
Pusa Road,  
New Delhi - 60

M/s Sandeep Instruments and Chemicals  
3229, Ranjeet Nagar  
New Deljhi-110008

M/s Fortune Biotech  
F-117, Katwaria Sarai  
ND-16

M/S Andromeda Scientific Corporation  
Plot No 584, Block-F,  
SGM Nagar, NIT  
Faridabad-121001